



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/755,932

01/13/2004

Steve T. Charles

225929

3455

23460

7590

02/25/2008

LEYDIG VOIT & MAYER, LTD  
TWO PRUDENTIAL PLAZA, SUITE 4900  
180 NORTH STETSON AVENUE  
CHICAGO, IL 60601-6731

EXAMINER

PHILOGENE, PEDRO

ART UNIT

PAPER NUMBER

3733

MAIL DATE

DELIVERY MODE

02/25/2008

PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

**Office Action Summary****Application No.**

10/755,932

**Applicant(s)**

CHARLES ET AL.

**Examiner**

Pedro Philogene

**Art Unit**

3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 30 November 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-8 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-8 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-946)
- 3) ☐ Information Disclosure Statement(s) (PTO/SE/US)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-8 are rejected under 35 U.S.C. 103(a) as being unpatentable over Furst et al. (6,245,028) in view of Gilboa (6,380,732).

With respect to the claims, Furst et al disclose a manipulator for use in performing medical procedures on a portion of a body of a patient, comprising: a medical tool (needle biopsy 210) a positioning mechanism which carries the medical tool, as best seen in FIG.3, the positioning mechanism being capable of moving the tool with at least one degree of freedom; a remote positioning tracking system; as best seen in FIG.4, capable of tracking the position of the medical tool relative to a predetermined frame of reference in three dimensions; as set forth in column 7, lines 20-67; columns 8-10, lines 1-67; the positioning mechanism comprises first and second arms which is pivotably support the medical tool and which are movable independently of each other; as best seen in FIG.3; further including a haptic interface which communicates with the manipulator controller that directs movement of the positioning mechanism, the haptic interface is adapted to receive manually input position information from an operator and communicate position signal based on the information to the manipulator controller; as set forth in column 7, lines 20-67; columns 8-10, lines 1-67.

It is noted that Furst et al did not teach of a position tracking system including an electromagnetic field generator for producing a multidimensional electromagnetic field in three mutual distinguishable dimensions the position tracking system being capable of tracking the position of the medical tool relative to a predetermined frame of reference in three-dimensions as the tool is moved by the positioning mechanism, as claimed by applicant. However, in a similar art, Gilboa, evidences the use of a position tracking system including an electromagnetic field generator for producing a multidimensional electromagnetic field in three mutual distinguishable dimensions to determine the three-dimensional position and three-axial rotational orientation of an object and to track a single object or a multiple objects simultaneously in a single system.

Therefore, given the teaching of Gilboa, it would have been obvious to one having ordinary skill in the art at the time the invention was made to use a position tracking system in the device of Furst et al, as taught by Gilboa to determine the three-dimensional position and three-axial rotational orientation of an object and to track translation and rotation of a single object or a multiple objects simultaneously in a single system.

#### ***Response to Amendment***

Applicant's arguments, see Remarks, filed 11/30/07, with respect to the rejection(s) of claim(s) 1-4 under 103 have been fully considered and are persuasive. Therefore, the rejection has been withdrawn. However, upon further consideration, a new ground(s) of rejection is made in view of Furst et al. Furst et al disclose a remote positioning tracking system capable of tracking the position of its needle biopsy.

***Conclusion***

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Pedro Philogene whose telephone number is (571) 272-4716. The examiner can normally be reached on Monday to Friday 6:30 AM to 4:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on (571) 272 - 4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3733

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Pedro Philogene/  
Primary Examiner, Art Unit 3733  
February 16, 2008